

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025490.D
 Acq On : 03 Dec 2021 18:16
 Operator : JC/MD
 Sample : M4924-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20211202

Quant Time: Dec 04 06:52:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	103809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177708	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62155	56.803	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.600%
35) Dibromofluoromethane	5.391	113	54952	50.398	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.653	98	206231	50.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.085	95	72413	46.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	25135	26.018	ug/l	99
12) 1,1-Dichloroethene	2.318	96	6351	6.601	ug/l	86
16) Acetone	2.379	43	2097	3.449	ug/l	96
24) 1,1-Dichloroethane	3.611	63	2148	1.185	ug/l #	95
27) cis-1,2-Dichloroethene	4.495	96	2840	2.307	ug/l	85
30) Chloroform	5.098	83	1536	0.815	ug/l	93
38) Carbon Tetrachloride	5.678	117	1624	0.951	ug/l	92
44) Trichloroethene	7.135	130	1180555	865.260	ug/l	97
55) 1,1,2-Trichloroethane	9.152	97	761	0.624	ug/l	91
64) Tetrachloroethene	9.274	164	2938	2.201	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025490.D
Acq On : 03 Dec 2021 18:16
Operator : JC/MD
Sample : M4924-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20211202

Quant Time: Dec 04 06:52:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

