

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120321\  
 Data File : VX025491.D  
 Acq On : 03 Dec 2021 18:39  
 Operator : JC/MD  
 Sample : M4924-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE135D1-20211202

Quant Time: Dec 04 06:53:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 17 15:36:16 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 102673   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 175103   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 156412   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 67514    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 62173    | 57.448   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 114.900% |
| 35) Dibromofluoromethane    | 5.391  | 113   | 54115    | 50.369   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 100.740% |
| 50) Toluene-d8              | 8.653  | 98    | 203653   | 50.415   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 100.840% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 69792    | 45.670   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 91.340%  |
| Target Compounds            |        |       |          |          |       |          |
|                             |        |       |          |          |       | Qvalue   |
| 30) Chloroform              | 5.098  | 83    | 4231     | 2.269    | ug/l  | 98       |
| 44) Trichloroethene         | 7.129  | 130   | 3032     | 2.255    | ug/l  | 86       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120321\  
Data File : VX025491.D  
Acq On : 03 Dec 2021 18:39  
Operator : JC/MD  
Sample : M4924-04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE135D1-20211202

Quant Time: Dec 04 06:53:09 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111721W.M  
Quant Title : SW846 8260  
QLast Update : Wed Nov 17 15:36:16 2021  
Response via : Initial Calibration

