

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025492.D
 Acq On : 03 Dec 2021 19:03
 Operator : JC/MD
 Sample : M4924-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D2-20211202

Quant Time: Dec 04 06:53:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177444	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62331	57.079	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.160%
35) Dibromofluoromethane	5.385	113	54949	50.470	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.940%
50) Toluene-d8	8.653	98	204515	49.961	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	69745	45.037	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%
Target Compounds						Qvalue
16) Acetone	2.380	43	1920	3.164	ug/l	97
44) Trichloroethene	7.135	130	1194	0.876	ug/l	84
45) 1,2-Dichloropropane	7.434	63	2287	1.986	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025492.D
Acq On : 03 Dec 2021 19:03
Operator : JC/MD
Sample : M4924-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D2-20211202

Quant Time: Dec 04 06:53:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

