

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025493.D
 Acq On : 03 Dec 2021 19:26
 Operator : JC/MD
 Sample : M4924-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20211202

Quant Time: Dec 04 06:53:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	173347	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60904	57.518	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 115.040%	
35) Dibromofluoromethane	5.391	113	53283	50.097	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.200%	
50) Toluene-d8	8.653	98	199825	49.969	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.940%	
62) 4-Bromofluorobenzene	11.079	95	70560	46.640	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.280%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	6828	7.304	ug/l	99
12) 1,1-Dichloroethene	2.319	96	504	0.541	ug/l #	70
16) Acetone	2.380	43	2205	3.748	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	3997	3.356	ug/l	85
44) Trichloroethene	7.129	130	19545	14.685	ug/l	90
45) 1,2-Dichloropropane	7.434	63	4463	3.966	ug/l	95
64) Tetrachloroethene	9.275	164	29724	23.021	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025493.D
Acq On : 03 Dec 2021 19:26
Operator : JC/MD
Sample : M4924-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20211202

Quant Time: Dec 04 06:53:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

