

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025494.D
 Acq On : 03 Dec 2021 19:50
 Operator : JC/MD
 Sample : M4924-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20211202

Quant Time: Dec 04 06:54:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61497	56.349	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.700%	
35) Dibromofluoromethane	5.391	113	54600	49.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.760%	
50) Toluene-d8	8.653	98	204043	49.575	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.160%	
62) 4-Bromofluorobenzene	11.085	95	70483	45.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.540%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	208924	216.827	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1394	1.453	ug/l #	71
16) Acetone	2.380	43	2221	3.662	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	928	0.756	ug/l	88
44) Trichloroethene	7.129	130	26502	19.347	ug/l	95
64) Tetrachloroethene	9.275	164	12902	9.848	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025494.D
Acq On : 03 Dec 2021 19:50
Operator : JC/MD
Sample : M4924-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20211202

Quant Time: Dec 04 06:54:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

