

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025503.D
 Acq On : 04 Dec 2021 00:07
 Operator : JC/MD
 Sample : M4923-01DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20211130DL

Quant Time: Dec 04 11:05:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176505	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61433	55.561	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.120%	
35) Dibromofluoromethane	5.385	113	54266	50.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.220%	
50) Toluene-d8	8.653	98	203887	50.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.140%	
62) 4-Bromofluorobenzene	11.079	95	71107	46.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.320%	
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	44235	32.642	ug/l	95
64) Tetrachloroethene	9.281	164	332	0.254	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025503.D
Acq On : 04 Dec 2021 00:07
Operator : JC/MD
Sample : M4923-01DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20211130DL

Quant Time: Dec 04 11:05:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

