

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025507.D
 Acq On : 04 Dec 2021 01:40
 Operator : JC/MD
 Sample : M4923-12DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20211201DL

Quant Time: Dec 04 11:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63569	56.920	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.840%
35) Dibromofluoromethane	5.391	113	56818	50.946	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.900%
50) Toluene-d8	8.653	98	208321	49.680	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	72535	45.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	672	0.682	ug/l	89
12) 1,1-Dichloroethene	2.325	96	309	0.315	ug/l #	81
44) Trichloroethene	7.129	130	48032	34.418	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025507.D
Acq On : 04 Dec 2021 01:40
Operator : JC/MD
Sample : M4923-12DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20211201DL

Quant Time: Dec 04 11:08:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

