

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX120321\
 Data File : VX025497.D
 Acq On : 03 Dec 2021 21:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 06:54:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	63	0.00
2 T	Dichlorodifluoromethane	0.433	0.443	-2.3	57	0.00
3 P	Chloromethane	0.537	0.457	14.9	52	0.00
4 C	Vinyl Chloride	0.589	0.505	14.3#	51	0.00
5 T	Bromomethane	0.253	0.263	-4.0	65	0.00
6 T	Chloroethane	0.407	0.353	13.3	57	0.00
7 T	Trichlorofluoromethane	0.875	0.832	4.9	58	0.00
8 T	Diethyl Ether	0.361	0.336	6.9	59	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.465	0.479	-3.0	65	0.00
10 T	Methyl Iodide	0.582	0.490	15.8	49#	0.00
11 T	Tert butyl alcohol	0.142	0.120	15.5	56	0.00
12 CM	1,1-Dichloroethene	0.463	0.451	2.6#	61	0.00
13 T	Acrolein	0.024	0.025	-4.2	62	0.00
14 T	Allyl chloride	0.839	0.751	10.5	59	0.00
15 T	Acrylonitrile	0.283	0.286	-1.1	63	0.00
16 T	Acetone	0.293	0.241	17.7	52	0.00
17 T	Carbon Disulfide	1.304	1.007	22.8	50	0.00
18 T	Methyl Acetate	0.936	0.902	3.6	63	0.00
19 T	Methyl tert-butyl Ether	1.615	1.637	-1.4	65	0.00
20 T	Methylene Chloride	0.551	0.524	4.9	65	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.496	2.4	63	0.00
22 T	Diisopropyl ether	1.597	1.602	-0.3	64	0.00
23 T	Vinyl Acetate	1.283	1.265	1.4	62	0.00
24 P	1,1-Dichloroethane	0.873	0.889	-1.8	65	0.00
25 T	2-Butanone	0.423	0.395	6.6	58	0.00
26 T	2,2-Dichloropropane	0.791	0.689	12.9	55	0.00
27 T	cis-1,2-Dichloroethene	0.593	0.603	-1.7	65	0.00
28 T	Bromochloromethane	0.407	0.422	-3.7	65	0.00
29 T	Tetrahydrofuran	0.262	0.255	2.7	62	0.00
30 C	Chloroform	0.908	0.944	-4.0#	66	0.00
31 T	Cyclohexane	0.814	0.755	7.2	59	0.00
32 T	1,1,1-Trichloroethane	0.832	0.834	-0.2	64	0.00
33 S	1,2-Dichloroethane-d4	0.527	0.554	-5.1	67	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
35 S	Dibromofluoromethane	0.307	0.314	-2.3	67	0.00
36 T	1,1-Dichloropropene	0.436	0.420	3.7	63	0.00
37 T	Ethyl Acetate	0.496	0.475	4.2	63	0.00
38 T	Carbon Tetrachloride	0.480	0.473	1.5	63	0.00
39 T	Methylcyclohexane	0.560	0.508	9.3	58	0.00
40 TM	Benzene	1.321	1.273	3.6	63	0.00
41 T	Methacrylonitrile	0.275	0.263	4.4	65	0.00
42 TM	1,2-Dichloroethane	0.430	0.459	-6.7	69	0.00
43 T	Isopropyl Acetate	0.826	0.783	5.2	62	0.00
44 TM	Trichloroethene	0.384	0.373	2.9	63	0.00
45 C	1,2-Dichloropropane	0.325	0.323	0.6#	64	0.00
46 T	Dibromomethane	0.236	0.238	-0.8	65	0.00
47 T	Bromodichloromethane	0.463	0.459	0.9	64	0.00
48 T	Methyl methacrylate	0.388	0.363	6.4	62	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
 Data File : VX025497.D
 Acq On : 03 Dec 2021 21:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 06:54:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	61	0.00
50 S	Toluene-d8	1.153	1.146	0.6	64	0.00
51 T	4-Methyl-2-Pentanone	0.518	0.513	1.0	64	0.00
52 CM	Toluene	0.863	0.843	2.3#	63	0.00
53 T	t-1,3-Dichloropropene	0.515	0.470	8.7	59	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.516	7.5	59	0.00
55 T	1,1,2-Trichloroethane	0.343	0.354	-3.2	66	0.00
56 T	Ethyl methacrylate	0.556	0.542	2.5	63	0.00
57 T	1,3-Dichloropropane	0.568	0.577	-1.6	66	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.270	-4.2	69	0.00
59 T	2-Hexanone	0.408	0.394	3.4	61	0.00
60 T	Dibromochloromethane	0.395	0.377	4.6	62	0.00
61 T	1,2-Dibromoethane	0.371	0.379	-2.2	66	0.00
62 S	4-Bromofluorobenzene	0.436	0.439	-0.7	67	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	66	0.00
64 T	Tetrachloroethene	0.412	0.396	3.9	63	0.00
65 PM	Chlorobenzene	1.019	1.009	1.0	66	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.378	3.3	64	0.00
67 C	Ethyl Benzene	1.761	1.721	2.3#	64	0.00
68 T	m/p-Xylenes	0.719	0.693	3.6	64	0.00
69 T	o-Xylene	0.704	0.692	1.7	65	0.00
70 T	Styrene	1.163	1.154	0.8	65	0.00
71 P	Bromoform	0.339	0.303	10.6	58	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
73 T	Isopropylbenzene	3.563	3.453	3.1	65	0.00
74 T	N-amyl acetate	1.541	1.445	6.2	63	0.00
75 P	1,1,2,2-Tetrachloroethane	1.153	1.131	1.9	66	0.00
76 T	1,2,3-Trichloropropane	1.024	1.001	2.2	64	0.00
77 T	Bromobenzene	0.916	0.912	0.4	66	0.00
78 T	n-propylbenzene	4.000	3.926	1.8	64	0.00
79 T	2-Chlorotoluene	2.435	2.359	3.1	66	0.00
80 T	1,3,5-Trimethylbenzene	2.995	2.891	3.5	65	0.00
81 T	trans-1,4-Dichloro-2-butene	0.406	0.313	22.9	50	0.00
82 T	4-Chlorotoluene	2.780	2.681	3.6	65	0.00
83 T	tert-Butylbenzene	3.044	2.944	3.3	66	0.00
84 T	1,2,4-Trimethylbenzene	2.960	2.905	1.9	65	0.00
85 T	sec-Butylbenzene	3.701	3.654	1.3	65	0.00
86 T	p-Isopropyltoluene	3.180	3.090	2.8	64	0.00
87 T	1,3-Dichlorobenzene	1.706	1.677	1.7	65	0.00
88 T	1,4-Dichlorobenzene	1.729	1.664	3.8	65	0.00
89 T	n-Butylbenzene	2.680	2.542	5.1	63	0.00
90 T	Hexachloroethane	0.586	0.495	15.5	56	0.00
91 T	1,2-Dichlorobenzene	1.663	1.618	2.7	65	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.241	0.233	3.3	61	0.00
93 T	1,2,4-Trichlorobenzene	0.979	1.010	-3.2	62	0.00
94 T	Hexachlorobutadiene	0.470	0.443	5.7	64	0.00
95 T	Naphthalene	3.046	3.374	-10.8	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120321\
Data File : VX025497.D
Acq On : 03 Dec 2021 21:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 04 06:54:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.957	1.020	-6.6	65	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6