

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX120418\
 Data File : VX006270.D
 Acq On : 04 Dec 2018 11:26
 Operator : JC/MD
 Sample : VX1204MBL01
 Misc : 5.0g/10mL/100ul/5.00ml/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1204MBL01

Quant Time: Dec 04 13:14:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	888568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1346112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1219441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	591835	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	447774	49.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.72%	
35) Dibromofluoromethane	5.50	113	430510	49.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.80%	
50) Toluene-d8	8.71	98	1554161	49.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.04%	
62) 4-Bromofluorobenzene	11.13	95	576199	48.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX120418\
Data File : VX006270.D
Acq On : 04 Dec 2018 11:26
Operator : JC/MD
Sample : VX1204MBL01
Misc : 5.0g/10mL/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1204MBL01

Quant Time: Dec 04 13:14:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

