

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX120418\
 Data File : VX006284.D
 Acq On : 04 Dec 2018 17:48
 Operator : JC/MD
 Sample : J6232-02DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0644DL

Quant Time: Dec 05 01:12:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	834614	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1247062	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1144653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	565310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	407134	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
35) Dibromofluoromethane	5.50	113	396909	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	8.71	98	1469411	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.13	95	539779	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
Target Compounds						
4) Vinyl Chloride	1.41	62	5389	0.596	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	69514	8.637	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	187426	20.878	ug/l	88
44) Trichloroethene	7.21	130	517692	54.331	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX120418\
Data File : VX006284.D
Acq On : 04 Dec 2018 17:48
Operator : JC/MD
Sample : J6232-02DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S32-0644DL

Quant Time: Dec 05 01:12:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

