

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120419\
 Data File : VX013750.D
 Acq On : 04 Dec 2019 15:30
 Operator : JC/SP
 Sample : K6095-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Dec 05 02:15:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	693350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1039509	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	924907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	408949	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	373869	46.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.90%	
35) Dibromofluoromethane	5.49	113	312071	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
50) Toluene-d8	8.71	98	1224490	48.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	412293	45.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120419\
Data File : VX013750.D
Acq On : 04 Dec 2019 15:30
Operator : JC/SP
Sample : K6095-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Dec 05 02:15:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

