

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120419\
 Data File : VX013751.D
 Acq On : 04 Dec 2019 15:54
 Operator : JC/SP
 Sample : K6129-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-191203

Quant Time: Dec 05 02:27:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	704006	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1056101	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	945744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	427927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	378771	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
35) Dibromofluoromethane	5.49	113	313993	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	8.71	98	1244111	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	422539	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
Target Compounds						
16) Acetone	2.43	43	27354	6.780	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120419\
Data File : VX013751.D
Acq On : 04 Dec 2019 15:54
Operator : JC/SP
Sample : K6129-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-191203

Quant Time: Dec 05 02:27:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

