

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120419\
 Data File : VX013758.D
 Acq On : 04 Dec 2019 18:37
 Operator : JC/SP
 Sample : K6098-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWL-C5-GW

Quant Time: Dec 05 03:25:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	695487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1045488	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	944189	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	441755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	373109	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
35) Dibromofluoromethane	5.49	113	315503	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	1233985	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	11.14	95	426592	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
Target Compounds						
					Qvalue	
16) Acetone	2.43	43	91089	22.854	ug/l	98
17) Carbon Disulfide	2.56	76	14589	0.773	ug/l	98
25) 2-Butanone	4.67	43	23840	4.025	ug/l	96
73) Isopropylbenzene	11.01	105	62166	2.006	ug/l	98
78) n-propylbenzene	11.35	91	66715	1.929	ug/l	99
83) tert-Butylbenzene	11.76	119	6305	0.244	ug/l	92
85) sec-Butylbenzene	11.94	105	55365	1.878	ug/l	86
89) n-Butylbenzene	12.38	91	16372	0.700	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120419\
Data File : VX013758.D
Acq On : 04 Dec 2019 18:37
Operator : JC/SP
Sample : K6098-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWL-C5-GW

Quant Time: Dec 05 03:25:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

