

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120419\
 Data File : VX013762.D
 Acq On : 04 Dec 2019 20:10
 Operator : JC/SP
 Sample : K6129-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-191203

Quant Time: Dec 05 05:57:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	678241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1031587	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	917433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	419589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	373100	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
35) Dibromofluoromethane	5.49	113	308219	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	1209105	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
62) 4-Bromofluorobenzene	11.14	95	412183	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	30658	3.797	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120419\
Data File : VX013762.D
Acq On : 04 Dec 2019 20:10
Operator : JC/SP
Sample : K6129-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-191203

Quant Time: Dec 05 05:57:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

