

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019716.D
 Acq On : 04 Dec 2020 13:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 04 16:13:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	355749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	574308	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	523096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	239481	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	212494	45.45	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	90.90%	
35) Dibromofluoromethane	5.46	113	173402	46.72	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	93.44%	
50) Toluene-d8	8.70	98	696302	49.14	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.12	95	249511	46.26	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	92.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	2.82	84	1768	Below Cal	#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019716.D
Acq On : 04 Dec 2020 13:37
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 04 16:13:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

