

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019721.D
 Acq On : 04 Dec 2020 16:58
 Operator : JC/MD
 Sample : VX1204WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1204WBL01

Quant Time: Dec 05 04:20:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	375353	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	602567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	528508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	227858	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	221332	44.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.74%
35) Dibromofluoromethane	5.46	113	181063	46.49	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.98%
50) Toluene-d8	8.70	98	713598	48.00	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.00%
62) 4-Bromofluorobenzene	11.12	95	245182	43.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.66%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019721.D
Acq On : 04 Dec 2020 16:58
Operator : JC/MD
Sample : VX1204WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1204WBL01

Quant Time: Dec 05 04:20:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

