

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019724.D
 Acq On : 04 Dec 2020 18:07
 Operator : JC/MD
 Sample : L4959-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF

Quant Time: Dec 05 04:25:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	333025	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	548006	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	484036	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	207599	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	204400	46.70	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.40%		
35) Dibromofluoromethane	5.46	113	166019	46.88	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.76%		
50) Toluene-d8	8.70	98	648721	47.98	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.96%		
62) 4-Bromofluorobenzene	11.12	95	226383	43.99	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.98%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019724.D
Acq On : 04 Dec 2020 18:07
Operator : JC/MD
Sample : L4959-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF

Quant Time: Dec 05 04:25:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

