

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019725.D
 Acq On : 04 Dec 2020 18:31
 Operator : JC/MD
 Sample : L4959-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 05 04:27:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	336602	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	548958	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	501462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	221025	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	209424	47.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.68%
35) Dibromofluoromethane	5.46	113	168112	47.38	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.76%
50) Toluene-d8	8.70	98	662313	48.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.80%
62) 4-Bromofluorobenzene	11.12	95	232360	45.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.14%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019725.D
Acq On : 04 Dec 2020 18:31
Operator : JC/MD
Sample : L4959-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 05 04:27:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

