

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019729.D
 Acq On : 04 Dec 2020 20:04
 Operator : JC/MD
 Sample : L4960-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Dec 05 04:43:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	335781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	549353	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	492090	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214012	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	206131	46.71	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.42%		
35) Dibromofluoromethane	5.46	113	166395	46.87	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.74%		
50) Toluene-d8	8.70	98	657883	48.54	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.08%		
62) 4-Bromofluorobenzene	11.12	95	227970	44.19	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.38%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019729.D
Acq On : 04 Dec 2020 20:04
Operator : JC/MD
Sample : L4960-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Dec 05 04:43:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

