

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019739.D
 Acq On : 04 Dec 2020 23:57
 Operator : JC/MD
 Sample : L4953-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ID039-FD01-120220

Quant Time: Dec 05 05:45:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	341849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	562792	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	513180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	229955	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	212558	47.31	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	94.62%	
35) Dibromofluoromethane	5.46	113	171202	47.07	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	94.14%	
50) Toluene-d8	8.70	98	679426	48.93	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	97.86%	
62) 4-Bromofluorobenzene	11.12	95	244060	46.18	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	92.36%	

Target Compounds					Qvalue
16) Acetone	2.42	43	2225	1.318 ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019739.D
Acq On : 04 Dec 2020 23:57
Operator : JC/MD
Sample : L4953-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ID039-FD01-120220

Quant Time: Dec 05 05:45:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

