

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120420\
 Data File : VX019741.D
 Acq On : 05 Dec 2020 00:43
 Operator : JC/MD
 Sample : L4953-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ID039-39MWE2S-120320

Quant Time: Dec 05 05:55:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	332423	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	547793	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	493422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214827	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	206739	47.32	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.64%	
35) Dibromofluoromethane		5.46	113	167330	47.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.52%	
50) Toluene-d8		8.70	98	653320	48.34	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.68%	
62) 4-Bromofluorobenzene		11.12	95	231931	45.08	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.31	50	1658	0.463	ug/l	96
16) Acetone	2.42	43	4857	2.958	ug/l	99
21) trans-1,2-Dichloroethene	3.14	96	2560	0.721	ug/l #	86
27) cis-1,2-Dichloroethene	4.56	96	59033	14.206	ug/l	90
44) Trichloroethene	7.18	130	2937	0.740	ug/l	100
51) 4-Methyl-2-Pentanone	8.62	43	1819	0.367	ug/l	97
87) 1,3-Dichlorobenzene	12.01	146	3532	0.512	ug/l	96
91) 1,2-Dichlorobenzene	12.38	146	47899	7.128	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120420\
Data File : VX019741.D
Acq On : 05 Dec 2020 00:43
Operator : JC/MD
Sample : L4953-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ID039-39MWE2S-120320

Quant Time: Dec 05 05:55:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

