

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX120420\
 Data File : VX019718.D
 Acq On : 04 Dec 2020 15:41
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX120420

Quant Time: Dec 05 04:12:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	50.329	-0.7	99	0.00
3 P	Chloromethane	50.000	45.099	9.8	97	0.00
4 C	Vinyl Chloride	50.000	49.250	1.5#	100	0.00
5 T	Bromomethane	50.000	49.486	1.0	101	0.00
6 T	Chloroethane	50.000	50.939	-1.9	101	0.00
7 T	Trichlorofluoromethane	50.000	49.231	1.5	101	0.00
8 T	Diethyl Ether	50.000	48.398	3.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.556	0.9	102	0.00
10 T	Methyl Iodide	50.000	49.841	0.3	102	0.00
11 T	Tert butyl alcohol	250.000	242.875	2.9	103	0.00
12 CM	1,1-Dichloroethene	50.000	48.798	2.4#	100	0.00
13 T	Acrolein	250.000	218.918	12.4	96	0.00
14 T	Allyl chloride	50.000	48.460	3.1	101	0.00
15 T	Acrylonitrile	250.000	237.665	4.9	100	0.00
16 T	Acetone	250.000	233.462	6.6	97	0.00
17 T	Carbon Disulfide	50.000	47.592	4.8	100	0.00
18 T	Methyl Acetate	50.000	47.890	4.2	99	0.00
19 T	Methyl tert-butyl Ether	50.000	49.685	0.6	102	0.00
20 T	Methylene Chloride	50.000	48.637	2.7	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.148	1.7	101	0.00
22 T	Diisopropyl ether	50.000	49.490	1.0	101	0.00
23 T	Vinyl Acetate	250.000	251.525	-0.6	102	0.00
24 P	1,1-Dichloroethane	50.000	48.772	2.5	101	0.00
25 T	2-Butanone	250.000	239.434	4.2	100	0.00
26 T	2,2-Dichloropropane	50.000	49.940	0.1	103	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.314	3.4	101	0.00
28 T	Bromochloromethane	50.000	48.862	2.3	102	0.00
29 T	Tetrahydrofuran	250.000	238.177	4.7	100	0.00
30 C	Chloroform	50.000	49.476	1.0#	102	0.00
31 T	Cyclohexane	50.000	49.640	0.7	100	0.00
32 T	1,1,1-Trichloroethane	50.000	49.211	1.6	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.969	6.1	104	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	48.033	3.9	103	0.00
36 T	1,1-Dichloropropene	50.000	49.480	1.0	102	0.00
37 T	Ethyl Acetate	50.000	47.448	5.1	101	0.00
38 T	Carbon Tetrachloride	50.000	50.244	-0.5	101	0.00
39 T	Methylcyclohexane	50.000	50.792	-1.6	102	0.00
40 TM	Benzene	50.000	48.820	2.4	101	0.00
41 T	Methacrylonitrile	50.000	48.422	3.2	101	0.00
42 TM	1,2-Dichloroethane	50.000	48.974	2.1	100	0.00
43 T	Isopropyl Acetate	50.000	48.490	3.0	102	0.00
44 TM	Trichloroethene	50.000	48.367	3.3	100	0.00
45 C	1,2-Dichloropropane	50.000	49.125	1.8#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX120420\
 Data File : VX019718.D
 Acq On : 04 Dec 2020 15:41
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX120420

Quant Time: Dec 05 04:12:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	49.289	1.4	101	0.00
47 T	Bromodichloromethane	50.000	49.908	0.2	101	0.00
48 T	Methyl methacrylate	50.000	49.042	1.9	101	0.00
49 T	1,4-Dioxane	1000.000	983.372	1.7	102	0.00
50 S	Toluene-d8	50.000	48.872	2.3	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	247.353	1.1	102	0.00
52 CM	Toluene	50.000	50.018	-0.0#	102	0.00
53 T	t-1,3-Dichloropropene	50.000	51.217	-2.4	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.106	-2.2	102	0.00
55 T	1,1,2-Trichloroethane	50.000	49.693	0.6	102	0.00
56 T	Ethyl methacrylate	50.000	51.121	-2.2	102	0.00
57 T	1,3-Dichloropropane	50.000	48.696	2.6	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	251.971	-0.8	104	0.00
59 T	2-Hexanone	250.000	251.892	-0.8	103	0.00
60 T	Dibromochloromethane	50.000	51.109	-2.2	102	0.00
61 T	1,2-Dibromoethane	50.000	49.935	0.1	104	0.00
62 S	4-Bromofluorobenzene	50.000	50.161	-0.3	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	49.081	1.8	101	0.00
65 PM	Chlorobenzene	50.000	49.028	1.9	101	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.070	-0.1	101	0.00
67 C	Ethyl Benzene	50.000	50.834	-1.7#	103	0.00
68 T	m/p-Xylenes	100.000	101.806	-1.8	102	0.00
69 T	o-Xylene	50.000	50.952	-1.9	103	0.00
70 T	Styrene	50.000	52.294	-4.6	103	0.00
71 P	Bromoform	50.000	53.100	-6.2	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	49.763	0.5	102	0.00
74 T	N-amyl acetate	50.000	50.303	-0.6	102	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.855	4.3	99	0.00
76 T	1,2,3-Trichloropropane	50.000	48.642	2.7	101	0.00
77 T	Bromobenzene	50.000	48.267	3.5	102	0.00
78 T	n-propylbenzene	50.000	50.462	-0.9	102	0.00
79 T	2-Chlorotoluene	50.000	49.121	1.8	102	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.537	-1.1	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.117	-4.2	104	0.00
82 T	4-Chlorotoluene	50.000	49.818	0.4	103	0.00
83 T	tert-Butylbenzene	50.000	50.622	-1.2	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.426	-2.9	102	0.00
85 T	sec-Butylbenzene	50.000	50.890	-1.8	102	0.00
86 T	p-Isopropyltoluene	50.000	52.081	-4.2	103	0.00
87 T	1,3-Dichlorobenzene	50.000	48.949	2.1	101	0.00
88 T	1,4-Dichlorobenzene	50.000	49.012	2.0	102	0.00
89 T	n-Butylbenzene	50.000	52.179	-4.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120420\
Data File : VX019718.D
Acq On : 04 Dec 2020 15:41
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX120420

Quant Time: Dec 05 04:12:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	50.629	-1.3	103	0.00
91 T	1,2-Dichlorobenzene	50.000	49.582	0.8	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.391	3.2	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.361	-0.7	101	0.00
94 T	Hexachlorobutadiene	50.000	49.269	1.5	104	0.00
95 T	Naphthalene	50.000	51.425	-2.8	102	0.00
96 T	1,2,3-Trichlorobenzene	50.000	50.189	-0.4	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6