

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120420\
 Data File : VX019744.D
 Acq On : 05 Dec 2020 01:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 06:52:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.458	0.472	-3.1	95	0.00
3 P	Chloromethane	0.538	0.483	10.2	91	0.00
4 C	Vinyl Chloride	0.568	0.587	-3.3#	98	0.00
5 T	Bromomethane	0.188	0.179	4.8	91	0.00
6 T	Chloroethane	0.341	0.367	-7.6	101	0.00
7 T	Trichlorofluoromethane	0.791	0.797	-0.8	97	0.00
8 T	Diethyl Ether	0.335	0.338	-0.9	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.474	-4.6	100	0.00
10 T	Methyl Iodide	0.668	0.652	2.4	94	0.00
11 T	Tert butyl alcohol	0.145	0.140	3.4	104	0.00
12 CM	1,1-Dichloroethene	0.458	0.487	-6.3#	102	0.00
13 T	Acrolein	0.073	0.074	-1.4	106	0.00
14 T	Allyl chloride	0.774	0.764	1.3	97	0.00
15 T	Acrylonitrile	0.297	0.305	-2.7	101	0.00
16 T	Acetone	0.247	0.241	2.4	95	0.00
17 T	Carbon Disulfide	1.314	1.287	2.1	96	0.00
18 T	Methyl Acetate	0.604	0.636	-5.3	102	0.00
19 T	Methyl tert-butyl Ether	1.668	1.741	-4.4	100	0.00
20 T	Methylene Chloride	0.639	0.601	5.9	101	0.00
21 T	trans-1,2-Dichloroethene	0.534	0.544	-1.9	98	0.00
22 T	Diisopropyl ether	1.564	1.619	-3.5	99	0.00
23 T	Vinyl Acetate	1.333	1.395	-4.7	99	0.00
24 P	1,1-Dichloroethane	0.943	0.961	-1.9	99	0.00
25 T	2-Butanone	0.380	0.390	-2.6	100	0.00
26 T	2,2-Dichloropropane	0.812	0.693	14.7	82	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.635	-1.6	100	0.00
28 T	Bromochloromethane	0.450	0.455	-1.1	99	0.00
29 T	Tetrahydrofuran	0.245	0.255	-4.1	102	0.00
30 C	Chloroform	0.970	0.996	-2.7#	99	0.00
31 T	Cyclohexane	0.817	0.835	-2.2	97	0.00
32 T	1,1,1-Trichloroethane	0.835	0.868	-4.0	99	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.645	1.8	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.323	0.317	1.9	101	0.00
36 T	1,1-Dichloropropene	0.452	0.443	2.0	98	0.00
37 T	Ethyl Acetate	0.456	0.460	-0.9	103	0.00
38 T	Carbon Tetrachloride	0.442	0.450	-1.8	99	0.00
39 T	Methylcyclohexane	0.531	0.519	2.3	94	0.00
40 TM	Benzene	1.380	1.380	0.0	99	0.00
41 T	Methacrylonitrile	0.250	0.249	0.4	99	0.00
42 TM	1,2-Dichloroethane	0.472	0.474	-0.4	99	0.00
43 T	Isopropyl Acetate	0.742	0.746	-0.5	101	0.00
44 TM	Trichloroethene	0.362	0.357	1.4	98	0.00
45 C	1,2-Dichloropropane	0.345	0.346	-0.3#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120420\
 Data File : VX019744.D
 Acq On : 05 Dec 2020 01:53
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 06:52:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.233	0.239	-2.6	101	0.00
47 T	Bromodichloromethane	0.462	0.472	-2.2	99	0.00
48 T	Methyl methacrylate	0.365	0.374	-2.5	102	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	104	0.00
50 S	Toluene-d8	1.234	1.237	-0.2	101	0.00
51 T	4-Methyl-2-Pentanone	0.453	0.470	-3.8	103	0.00
52 CM	Toluene	0.863	0.869	-0.7#	99	0.00
53 T	t-1,3-Dichloropropene	0.513	0.517	-0.8	96	0.00
54 T	cis-1,3-Dichloropropene	0.557	0.561	-0.7	96	0.00
55 T	1,1,2-Trichloroethane	0.348	0.354	-1.7	101	0.00
56 T	Ethyl methacrylate	0.526	0.554	-5.3	101	0.00
57 T	1,3-Dichloropropane	0.594	0.597	-0.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.279	0.284	-1.8	101	0.00
59 T	2-Hexanone	0.351	0.362	-3.1	102	0.00
60 T	Dibromochloromethane	0.353	0.369	-4.5	100	0.00
61 T	1,2-Dibromoethane	0.369	0.372	-0.8	100	0.00
62 S	4-Bromofluorobenzene	0.470	0.465	1.1	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.335	0.328	2.1	97	0.00
65 PM	Chlorobenzene	0.996	0.987	0.9	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.371	-3.3	101	0.00
67 C	Ethyl Benzene	1.768	1.795	-1.5#	99	0.00
68 T	m/p-Xylenes	0.666	0.672	-0.9	98	0.00
69 T	o-Xylene	0.641	0.650	-1.4	98	0.00
70 T	Styrene	1.082	1.120	-3.5	98	0.00
71 P	Bromoform	0.282	0.295	-4.6	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.416	3.445	-0.8	99	0.00
74 T	N-amyl acetate	1.369	1.412	-3.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.204	1.222	-1.5	100	0.00
76 T	1,2,3-Trichloropropane	1.121	1.128	-0.6	99	0.00
77 T	Bromobenzene	0.874	0.864	1.1	99	0.00
78 T	n-propylbenzene	3.902	3.961	-1.5	98	0.00
79 T	2-Chlorotoluene	2.403	2.375	1.2	97	0.00
80 T	1,3,5-Trimethylbenzene	2.883	2.905	-0.8	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.388	0.383	1.3	93	0.00
82 T	4-Chlorotoluene	2.814	2.799	0.5	98	0.00
83 T	tert-Butylbenzene	2.828	2.865	-1.3	97	0.00
84 T	1,2,4-Trimethylbenzene	2.904	2.953	-1.7	96	0.00
85 T	sec-Butylbenzene	3.228	3.249	-0.7	96	0.00
86 T	p-Isopropyltoluene	2.959	2.976	-0.6	95	0.00
87 T	1,3-Dichlorobenzene	1.604	1.566	2.4	95	0.00
88 T	1,4-Dichlorobenzene	1.634	1.610	1.5	98	0.00
89 T	n-Butylbenzene	2.621	2.591	1.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX120420\
Data File : VX019744.D
Acq On : 05 Dec 2020 01:53
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 05 06:52:22 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.511	0.522	-2.2	99	0.00
91 T	1,2-Dichlorobenzene	1.564	1.551	0.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.259	0.270	-4.2	104	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.006	1.2	94	0.00
94 T	Hexachlorobutadiene	0.455	0.413	9.2	91	0.00
95 T	Naphthalene	3.301	3.533	-7.0	101	0.00
96 T	1,2,3-Trichlorobenzene	0.995	0.985	1.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6