

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025514.D
 Acq On : 04 Dec 2021 12:04
 Operator : JC/MD
 Sample : VX1204WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1204WBL01

Quant Time: Dec 06 01:00:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67484	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61448	55.876	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.760%
35) Dibromofluoromethane	5.391	113	54312	50.582	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	8.647	98	201077	49.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.079	95	68118	44.601	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025514.D
Acq On : 04 Dec 2021 12:04
Operator : JC/MD
Sample : VX1204WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1204WBL01

Quant Time: Dec 06 01:00:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

