

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025517.D
 Acq On : 04 Dec 2021 13:23
 Operator : JC/MD
 Sample : M4924-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20211202DL

Quant Time: Dec 06 01:01:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175950	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67273	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61419	55.45	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.90%
35) Dibromofluoromethane	5.391	113	54797	50.76	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.52%
50) Toluene-d8	8.653	98	201490	49.64	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.28%
62) 4-Bromofluorobenzene	11.085	95	69651	45.36	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.72%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	40685	41.60	ug/l	98
12) 1,1-Dichloroethene	2.325	96	351	0.36	ug/l #	58
44) Trichloroethene	7.129	130	5312	3.93	ug/l	98
64) Tetrachloroethene	9.275	164	2553	1.97	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025517.D
Acq On : 04 Dec 2021 13:23
Operator : JC/MD
Sample : M4924-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20211202DL

Quant Time: Dec 06 01:01:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

