

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025518.D
 Acq On : 04 Dec 2021 13:46
 Operator : JC/MD
 Sample : M4924-03DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20211202DL

Quant Time: Dec 06 01:01:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	103646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67703	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60671	55.53	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.06%	
35) Dibromofluoromethane	5.391	113	53426	50.48	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.96%	
50) Toluene-d8	8.653	98	197901	49.73	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.46%	
62) 4-Bromofluorobenzene	11.079	95	69329	46.05	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.10%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1902	1.97	ug/l	98
12) 1,1-Dichloroethene	2.319	96	347	0.36	ug/l	# 80
44) Trichloroethene	7.129	130	75326	56.88	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025518.D
Acq On : 04 Dec 2021 13:46
Operator : JC/MD
Sample : M4924-03DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20211202DL

Quant Time: Dec 06 01:01:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

