

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025521.D
 Acq On : 04 Dec 2021 14:57
 Operator : JC/MD
 Sample : M4923-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20211201

Quant Time: Dec 06 01:02:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	98180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	164025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58378	56.410	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.820%
35) Dibromofluoromethane	5.391	113	51208	50.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.760%
50) Toluene-d8	8.653	98	188312	49.766	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.085	95	65190	45.539	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.080%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1959	3.407	ug/l	99
44) Trichloroethene	7.129	130	575	0.457	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025521.D
Acq On : 04 Dec 2021 14:57
Operator : JC/MD
Sample : M4923-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20211201

Quant Time: Dec 06 01:02:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

