

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025522.D
 Acq On : 04 Dec 2021 15:20
 Operator : JC/MD
 Sample : M4923-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20211202

Quant Time: Dec 06 01:02:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	98282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	164689	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58671	56.635	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.260%
35) Dibromofluoromethane	5.391	113	51238	50.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.420%
50) Toluene-d8	8.653	98	190239	50.073	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.085	95	65822	45.796	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.600%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1060	1.841	ug/l	99
20) Methylene Chloride	2.794	84	383	0.354	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025522.D
Acq On : 04 Dec 2021 15:20
Operator : JC/MD
Sample : M4923-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20211202

Quant Time: Dec 06 01:02:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

