

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025524.D
 Acq On : 04 Dec 2021 16:07
 Operator : JC/MD
 Sample : M4924-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20211202

Quant Time: Dec 06 01:03:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	98947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	166953	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59167	56.729	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.460%	
35) Dibromofluoromethane	5.391	113	51908	50.673	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.340%	
50) Toluene-d8	8.653	98	193420	50.220	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.079	95	67607	46.400	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.800%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	26422	28.694	ug/l	98
12) 1,1-Dichloroethene	2.319	96	16192	17.656	ug/l	92
16) Acetone	2.380	43	1962	3.385	ug/l	93
24) 1,1-Dichloroethane	3.617	63	2792	1.617	ug/l	93
27) cis-1,2-Dichloroethene	4.495	96	2780	2.370	ug/l	87
30) Chloroform	5.105	83	1268	0.705	ug/l #	81
38) Carbon Tetrachloride	5.678	117	2141	1.335	ug/l	87
44) Trichloroethene	7.129	130	328514	256.287	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025524.D
Acq On : 04 Dec 2021 16:07
Operator : JC/MD
Sample : M4924-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20211202

Quant Time: Dec 06 01:03:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

