

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025531.D
 Acq On : 04 Dec 2021 18:51
 Operator : JC/MD
 Sample : M4923-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20211201

Quant Time: Dec 06 01:07:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	94667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	162085	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56829	56.951	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.900%
35) Dibromofluoromethane	5.391	113	49603	49.877	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.760%
50) Toluene-d8	8.653	98	186363	49.841	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.680%
62) 4-Bromofluorobenzene	11.085	95	66294	46.865	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.740%
Target Compounds						
16) Acetone	2.379	43	2220	4.004	ug/l	94
44) Trichloroethene	7.129	130	1215	0.976	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025531.D
Acq On : 04 Dec 2021 18:51
Operator : JC/MD
Sample : M4923-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20211201

Quant Time: Dec 06 01:07:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

