

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025533.D
 Acq On : 04 Dec 2021 19:38
 Operator : JC/MD
 Sample : M4923-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D-20211201

Quant Time: Dec 06 01:08:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61358	57.057	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.120%	
35) Dibromofluoromethane	5.391	113	52723	49.584	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.160%	
50) Toluene-d8	8.653	98	199725	49.957	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.920%	
62) 4-Bromofluorobenzene	11.079	95	72883	48.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.380%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	13270	13.977	ug/l	96
12) 1,1-Dichloroethene	2.319	96	2101	2.222	ug/l	100
16) Acetone	2.380	43	2333	3.904	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	3601	2.977	ug/l	86
44) Trichloroethene	7.129	130	130516	98.091	ug/l	94
64) Tetrachloroethene	9.275	164	823	0.630	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025533.D
Acq On : 04 Dec 2021 19:38
Operator : JC/MD
Sample : M4923-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D-20211201

Quant Time: Dec 06 01:08:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

