

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025534.D
 Acq On : 04 Dec 2021 20:02
 Operator : JC/MD
 Sample : M4923-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20211201

Quant Time: Dec 06 01:09:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163827	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58149	57.315	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.640%	
35) Dibromofluoromethane	5.391	113	50672	50.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.820%	
50) Toluene-d8	8.653	98	188815	49.959	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.920%	
62) 4-Bromofluorobenzene	11.085	95	68785	48.109	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.220%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	13820	15.429	ug/l	99
12) 1,1-Dichloroethene	2.319	96	4738	5.311	ug/l	92
16) Acetone	2.380	43	1484	2.632	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	2027	1.776	ug/l	80
30) Chloroform	5.111	83	1294	0.740	ug/l	99
38) Carbon Tetrachloride	5.684	117	1461	0.928	ug/l	89
44) Trichloroethene	7.129	130	250172	198.893	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
Data File : VX025534.D
Acq On : 04 Dec 2021 20:02
Operator : JC/MD
Sample : M4923-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20211201

Quant Time: Dec 06 01:09:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

