



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120421\
 Data File : VX025511.D
 Acq On : 04 Dec 2021 10:00
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Title : SW846 8260
 Last Update : Wed Nov 17 15:36:16 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	5111	PASS
75	95	30	60	50.2	14027	PASS
95	95	100	100	100.0	27957	PASS
96	95	5	9	7.2	2026	PASS
173	174	0.00	2	1.1	252	PASS
174	95	50	100	80.4	22464	PASS
175	174	5	9	7.7	1737	PASS
176	174	95	101	97.0	21786	PASS
177	176	5	9	6.9	1499	PASS