

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044116.D
 Acq On : 04 Dec 2024 14:56
 Operator : JC/MD
 Sample : P5006-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C10-1DL

Quant Time: Dec 05 00:27:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107028	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	209800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96351	52.487	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.980%			
35) Dibromofluoromethane	5.385	113	68065	45.403	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 90.800%			
50) Toluene-d8	8.647	98	258636	50.049	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.100%			
62) 4-Bromofluorobenzene	11.079	95	87724	49.880	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.760%			
Target Compounds						
16) Acetone	2.386	43	4639	5.502	ug/l #	87
27) cis-1,2-Dichloroethene	4.501	96	1929	1.226	ug/l	92
40) Benzene	6.037	78	27877	4.791	ug/l	98
52) Toluene	8.714	92	56542	16.234	ug/l	99
67) Ethyl Benzene	10.189	91	285080	42.010	ug/l	100
68) m/p-Xylenes	10.299	106	42056	16.619	ug/l	99
69) o-Xylene	10.640	106	31086	12.581	ug/l	94
73) Isopropylbenzene	10.963	105	12846	2.114	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	12227	2.446	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	43311	8.703	ug/l	98
95) Naphthalene	13.774	128	3511065	620.989	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044116.D
Acq On : 04 Dec 2024 14:56
Operator : JC/MD
Sample : P5006-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C10-1DL

Quant Time: Dec 05 00:27:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

