

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044118.D
 Acq On : 04 Dec 2024 15:43
 Operator : JC/MD
 Sample : P5006-10DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C10-2DL

Quant Time: Dec 05 00:28:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105677	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	201675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74917	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94308	52.031	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.060%
35) Dibromofluoromethane	5.379	113	66594	46.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.420%
50) Toluene-d8	8.647	98	248998	50.125	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	86307	51.052	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.100%

Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.483	96	1723	1.109	ug/l	93
40) Benzene	6.038	78	44054	7.876	ug/l	98
52) Toluene	8.714	92	51529	15.391	ug/l	98
67) Ethyl Benzene	10.189	91	185446	28.430	ug/l	99
68) m/p-Xylenes	10.299	106	26667	10.963	ug/l	98
69) o-Xylene	10.640	106	19428	8.180	ug/l	95
73) Isopropylbenzene	10.957	105	7004	1.216	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	6984	1.474	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	23088	4.895	ug/l	99
95) Naphthalene	13.774	128	1843051	343.935	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044118.D
Acq On : 04 Dec 2024 15:43
Operator : JC/MD
Sample : P5006-10DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C10-2DL

Quant Time: Dec 05 00:28:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

