

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044121.D
 Acq On : 04 Dec 2024 16:52
 Operator : JC/MD
 Sample : P5006-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C5-3RE

Quant Time: Dec 05 00:30:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105660	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	206619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	181269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94863	52.345	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.700%			
35) Dibromofluoromethane	5.385	113	51791	35.079	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 70.160%#			
50) Toluene-d8	8.647	98	253556	49.822	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.640%			
62) 4-Bromofluorobenzene	11.079	95	88136	50.886	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.780%			
Target Compounds						
16) Acetone	2.386	43	14667	17.621	ug/l	92
27) cis-1,2-Dichloroethene	4.483	96	1743	1.122	ug/l	95
40) Benzene	6.031	78	9135	1.594	ug/l	100
67) Ethyl Benzene	10.189	91	293251	43.470	ug/l	97
68) m/p-Xylenes	10.305	106	30513	12.129	ug/l	99
69) o-Xylene	10.640	106	18622	7.581	ug/l	97
73) Isopropylbenzene	10.963	105	34271	5.633	ug/l	97
78) n-propylbenzene	11.305	91	9457	1.392	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	26118	5.218	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	104947	21.060	ug/l	98
95) Naphthalene	13.774	128	6222910	1099.191	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044121.D
Acq On : 04 Dec 2024 16:52
Operator : JC/MD
Sample : P5006-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C5-3RE

Quant Time: Dec 05 00:30:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

