

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044122.D
 Acq On : 04 Dec 2024 17:15
 Operator : JC/MD
 Sample : P5006-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C6-1RE

Quant Time: Dec 05 00:31:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105046	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	206576	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96662	53.650	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.300%			
35) Dibromofluoromethane	5.385	113	50712	34.356	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 68.720%#			
50) Toluene-d8	8.647	98	248214	48.782	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.560%			
62) 4-Bromofluorobenzene	11.079	95	85139	49.166	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.340%			
Target Compounds						
16) Acetone	2.380	43	22057	26.655	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	5820	3.768	ug/l	94
67) Ethyl Benzene	10.195	91	14746	2.239	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	6386	1.397	ug/l	93
95) Naphthalene	13.774	128	488271	94.029	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044122.D
Acq On : 04 Dec 2024 17:15
Operator : JC/MD
Sample : P5006-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C6-1RE

Quant Time: Dec 05 00:31:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

