

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044123.D
 Acq On : 04 Dec 2024 17:38
 Operator : JC/MD
 Sample : P5006-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C6-3RE

Quant Time: Dec 05 00:32:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108195	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	212238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85975	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97519	52.550	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.100%		
35) Dibromofluoromethane	5.385	113	52592	34.679	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 69.360%#		
50) Toluene-d8	8.647	98	259809	49.699	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.400%		
62) 4-Bromofluorobenzene	11.079	95	92267	51.861	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.720%		
Target Compounds						
16) Acetone	2.386	43	17474	20.502	ug/l	Qvalue 93
27) cis-1,2-Dichloroethene	4.489	96	1632	1.026	ug/l	92
67) Ethyl Benzene	10.195	91	62249	8.961	ug/l	100
68) m/p-Xylenes	10.305	106	5664	2.187	ug/l	94
69) o-Xylene	10.640	106	5646	2.232	ug/l	97
73) Isopropylbenzene	10.957	105	8047	1.218	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	6975	1.283	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	29252	5.404	ug/l	99
95) Naphthalene	13.774	128	2128069	346.046	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044123.D
Acq On : 04 Dec 2024 17:38
Operator : JC/MD
Sample : P5006-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C6-3RE

Quant Time: Dec 05 00:32:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

