

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044129.D
 Acq On : 04 Dec 2024 19:57
 Operator : JC/MD
 Sample : P5006-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C5-2RE

Quant Time: Dec 05 00:37:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101936	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	199845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	169469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91520	52.346	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.700%			
35) Dibromofluoromethane	5.385	113	49951	34.980	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 69.960%#			
50) Toluene-d8	8.647	98	245054	49.783	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.560%			
62) 4-Bromofluorobenzene	11.079	95	80566	48.092	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.180%			
Target Compounds						
16) Acetone	2.380	43	24857	30.955	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	7330	4.890	ug/l	89
40) Benzene	6.037	78	14623	2.638	ug/l	97
52) Toluene	8.726	92	3847	1.160	ug/l	91
67) Ethyl Benzene	10.195	91	86422	13.703	ug/l	98
68) m/p-Xylenes	10.299	106	11205	4.764	ug/l	99
69) o-Xylene	10.640	106	8554	3.725	ug/l	99
73) Isopropylbenzene	10.963	105	8899	1.693	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	7833	1.811	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	30130	6.998	ug/l	94
95) Naphthalene	13.774	128	1759335	359.667	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044129.D
Acq On : 04 Dec 2024 19:57
Operator : JC/MD
Sample : P5006-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C5-2RE

Quant Time: Dec 05 00:37:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

