

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044130.D
 Acq On : 04 Dec 2024 20:21
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 05 00:37:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	217531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98907	52.008	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.020%			
35) Dibromofluoromethane	5.385	113	69886	44.961	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 89.920%			
50) Toluene-d8	8.647	98	267413	49.909	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.820%			
62) 4-Bromofluorobenzene	11.079	95	92498	50.726	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.460%			
Target Compounds						
95) Naphthalene	13.774	128	39544	6.650	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044130.D
Acq On : 04 Dec 2024 20:21
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 05 00:37:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

