

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044131.D
 Acq On : 04 Dec 2024 20:44
 Operator : JC/MD
 Sample : P5093-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LL-001

Quant Time: Dec 05 00:38:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	106401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	205400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95934	52.568	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.140%			
35) Dibromofluoromethane	5.385	113	67460	45.963	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.920%			
50) Toluene-d8	8.647	98	253866	50.178	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.360%			
62) 4-Bromofluorobenzene	11.079	95	84962	49.345	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.680%			
Target Compounds						
16) Acetone	2.380	43	1115	1.330	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044131.D
Acq On : 04 Dec 2024 20:44
Operator : JC/MD
Sample : P5093-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LL-001

Quant Time: Dec 05 00:38:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

