

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
 Data File : VX044132.D
 Acq On : 04 Dec 2024 21:07
 Operator : JC/MD
 Sample : P5093-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LL-001-FB-12-4-24

Quant Time: Dec 05 00:38:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	110604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218721	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101329	53.414	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.820%
35) Dibromofluoromethane	5.385	113	71404	45.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	8.647	98	262132	48.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	90619	49.425	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044132.D
Acq On : 04 Dec 2024 21:07
Operator : JC/MD
Sample : P5093-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LL-001-FB-12-4-24

Quant Time: Dec 05 00:38:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

