



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120424\
Data File : VX044108.D
Acq On : 04 Dec 2024 11:12
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Title : SW846 8260
Last Update : Fri Nov 22 03:45:52 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	7067	PASS
75	95	30	60	56.7	19328	PASS
95	95	100	100	100.0	34104	PASS
96	95	5	9	6.1	2087	PASS
173	174	0.00	2	1.0	232	PASS
174	95	50	100	68.2	23267	PASS
175	174	5	9	7.7	1801	PASS
176	174	95	101	95.1	22136	PASS
177	176	5	9	6.6	1450	PASS