

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006294.D
 Acq On : 05 Dec 2018 12:24
 Operator : JC/MD
 Sample : J6210-03ME
 Misc : 5.70g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3(4.5)ME

Quant Time: Dec 06 01:12:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	715605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1070788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	981738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	527220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	344962	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
35) Dibromofluoromethane	5.51	113	334040	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	8.71	98	1241878	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
62) 4-Bromofluorobenzene	11.14	95	484378	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
Target Compounds						
64) Tetrachloroethene	9.34	164	291154	39.380	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006294.D
Acq On : 05 Dec 2018 12:24
Operator : JC/MD
Sample : J6210-03ME
Misc : 5.70g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3(4.5)ME

Quant Time: Dec 06 01:12:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

