

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006295.D
 Acq On : 05 Dec 2018 12:51
 Operator : JC/MD
 Sample : J6210-05ME
 Misc : 6.22g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-9(4.5)ME

Quant Time: Dec 06 00:35:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	724707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1084313	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	991992	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	525853	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	353186	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
35) Dibromofluoromethane	5.50	113	339834	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	8.71	98	1249929	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
62) 4-Bromofluorobenzene	11.14	95	484985	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	23447	3.008	ug/l	87
44) Trichloroethene	7.21	130	141491	17.078	ug/l	96
64) Tetrachloroethene	9.34	164	689436	92.287	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006295.D
Acq On : 05 Dec 2018 12:51
Operator : JC/MD
Sample : J6210-05ME
Misc : 6.22g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-9(4.5)ME

Quant Time: Dec 06 00:35:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

