

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006297.D
 Acq On : 05 Dec 2018 13:44
 Operator : JC/MD
 Sample : J6214-33
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-4

Quant Time: Dec 06 00:41:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	841342	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1265711	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1141269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	554103	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	411462	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
35) Dibromofluoromethane	5.51	113	397709	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1476629	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.14	95	540861	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
30) Chloroform	5.22	83	18480	1.310	ug/l	87
64) Tetrachloroethene	9.33	164	20625	2.400	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006297.D
Acq On : 05 Dec 2018 13:44
Operator : JC/MD
Sample : J6214-33
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-4

Quant Time: Dec 06 00:41:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

