

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006300.D
 Acq On : 05 Dec 2018 15:04
 Operator : JC/MD
 Sample : J6241-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-181204

Quant Time: Dec 06 00:54:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	821559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1250641	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1128103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	542519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	407495	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
35) Dibromofluoromethane	5.51	113	389062	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
50) Toluene-d8	8.71	98	1450002	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	540087	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
Target Compounds						
16) Acetone	2.45	43	23092	5.867	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	42165	4.772	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006300.D
Acq On : 05 Dec 2018 15:04
Operator : JC/MD
Sample : J6241-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-181204

Quant Time: Dec 06 00:54:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

