

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006302.D
 Acq On : 05 Dec 2018 15:58
 Operator : JC/MD
 Sample : J6233-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-2

Quant Time: Dec 06 01:00:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	807541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1228193	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1121593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	542507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	408198	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	5.50	113	166270	20.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	41.82%	
50) Toluene-d8	8.71	98	1438975	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.14	95	535073	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
Target Compounds						
16) Acetone	2.45	43	221593	57.278	ug/l	99
20) Methylene Chloride	2.86	84	895196	108.725	ug/l	97
95) Naphthalene	13.83	128	73808	1.771	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006302.D
Acq On : 05 Dec 2018 15:58
Operator : JC/MD
Sample : J6233-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-2

Quant Time: Dec 06 01:00:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

