

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120518\
 Data File : VX006304.D
 Acq On : 05 Dec 2018 16:51
 Operator : JC/MD
 Sample : J6243-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-WATER-B

Quant Time: Dec 06 01:04:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	801033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1213594	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1106304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	548929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	402994	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
35) Dibromofluoromethane	5.51	113	384801	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
50) Toluene-d8	8.71	98	1424680	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	11.13	95	524304	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
Target Compounds						
11) Tert butyl alcohol	3.06	59	61106	27.325	ug/l	99
16) Acetone	2.45	43	224871	58.598	ug/l	96
25) 2-Butanone	4.70	43	46616	8.513	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120518\
Data File : VX006304.D
Acq On : 05 Dec 2018 16:51
Operator : JC/MD
Sample : J6243-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-WATER-B

Quant Time: Dec 06 01:04:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

