

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120519\
 Data File : VX013767.D
 Acq On : 05 Dec 2019 11:10
 Operator : JC/SP
 Sample : VX1205WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1205WBL01

Quant Time: Dec 06 02:40:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	111777	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	631890	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	536206	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.05	65	242370	29.36	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	97.87%
60) 4-Bromofluorobenzene	11.13	95	239280	27.76	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.53%
63) Toluene-d8	8.71	98	730198	30.11	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.37%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120519\
Data File : VX013767.D
Acq On : 05 Dec 2019 11:10
Operator : JC/SP
Sample : VX1205WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1205WBL01

Quant Time: Dec 06 02:40:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

